

*ClaroNav Kolahi Inc.
1140 Sheppard Ave. West, Unit 10 Toronto, ON Canada M3K 2A2
Tel. 1-647-951-1525 Fax 1-647-951-1524
www.claronav.com*

Valid as of July 4th, 2017

Declaration of Conformity

This declaration is made in accordance with Medical Device Directive 93/42/EEC, Annex II

Name and Address of the Manufacturer:

ClaroNav Kolahi Inc.
1140 Sheppard Ave. West, Unit 10
Toronto, Ontario, CANADA, M3K 2A2
Ahmad Kolahi - CEO - ahmad@claronav.com

Name and Address of Authorized Representative:

Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands
Tel: (31) (0) 70 345-8570
Fax: (31) (0) 70 346-7299
Web: <http://www.emergogroup.com/>

Identification of the Device: NaviENT System Model # 955-NE-NE

Name and Address of Notified Body:

BSI
Kitemark Court Davy Avenue Knowlhill
Milton Keynes MK5 8PP
United Kingdom
Phone: +44 (0) 8450 809000
Fax: +44 (0) 8450 809000
Email: product.certification@bsigroup.com

Notified Body number: 0086

Certificate Number (issued by BSI): CE 656684

Classification of the device: Class IIa (as per Rule 9)

**Compliance with Annex II (excluding section 4) of the Medical Device Directive,
Council Directive 93/42/EEC**

ClaroNav Kolahi Inc. declares that NaviENT bears the CE mark of conformity and complies with the requirement of council directive of 93/42/EEC for Medical Devices. All supporting documentation is retained at the premises of the manufacturer.

This Declaration of Conformity covers the products listed in Appendix A which bears the CE mark of Conformity and is distributed by ClaroNav Kolahi Inc. in accordance to the Medical Device Directive.

I, the undersigned, hereby declare that the products identified in Appendix A conforms to the above Medical Device Directive.



Company's Representative:

Signature:

A. Kolahi

Name: Ahmad Kolahi

Title: ClaroNav Kolahi Inc., CEO

Date: July 4th, 2017

Appendix A- NaviENT Product List

This product list specifies the NaviENT products bearing the CE Mark distributed by ClaroNav Kolahi Inc. in accordance to the EU Medical Device Directive:

Product Model	REF	Initial Serial Number
NaviENT System	955-NE-NE	1601